



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay
Jim Doyle, Governor
Richard J. Leinenkugel, Secretary

February 17, 2009

SIRONA DENTAL SYSTEMS, LLC
MOHAMAD HAMDAN
4835 SIRONA DR SUITE 100
CHARLOTTE NC 28273

SIRONA DENTAL SYSTEMS, BMBH
MOHAMAD HAMDAN
FABRIKSTRASSE 31
BENSHEIM D-64625
GERMANY

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: SIRONA DENTAL SYSTEMS, BMBH
Product Name: SIRONA C8+
Model Number(s): D3417
Product File No: 20090083

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of February 2014.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval supersedes the approval issued on July 19, 2004 under product file number 20040314.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20040314.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov